

Manual Supplement

Manual Title: PS410 Users
Part Number: 2631795
Print Date: April 2006
Revision/Date: 1, 12/07

Supplement Issue: 1
Issue Date: 7/08
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 5, under **ECG, Adult or Pediatrics Waveforms**, change:

From: ECG Amplitudes.....0.5, 1.0, 1.5 and 2.0 mV

To: ECG Amplitudes.....0.5, 1.0, and 2.0 mV